

## § 606.20

(1) The blood or blood components have left the control of the licensed manufacturer, unlicensed registered blood establishment, or transfusion service; or

(2) The licensed manufacturer has provided Source Plasma or any other blood component for use in the manufacture of a licensed biological product.

(1) *Control* means having responsibility for maintaining the continued safety, purity, and potency of the product and for compliance with applicable product and establishment standards, and for compliance with current good manufacturing practices.

[40 FR 53532, Nov. 18, 1975, as amended at 64 FR 45370, Aug. 19, 1999; 65 FR 66635, Nov. 7, 2000; 66 FR 1835, Jan. 10, 2001; 66 FR 40889, Aug. 6, 2001]

### Subpart B—Organization and Personnel

#### § 606.20 Personnel.

(a) [Reserved]

(b) The personnel responsible for the collection, processing, compatibility testing, storage or distribution of blood or blood components shall be adequate in number, educational background, training and experience, including professional training as necessary, or combination thereof, to assure competent performance of their assigned functions, and to ensure that the final product has the safety, purity, potency, identity and effectiveness it purports or is represented to possess. All personnel shall have capabilities commensurate with their assigned functions, a thorough understanding of the procedures or control operations they perform, the necessary training or experience, and adequate information concerning the application of pertinent provisions of this part to their respective functions.

(c) Persons whose presence can adversely affect the safety and purity of the products shall be excluded from areas where the collection, processing, compatibility testing, storage or distribution of blood or blood components is conducted.

[40 FR 53532, Nov. 18, 1975, as amended at 49 FR 23833, June 8, 1984; 55 FR 11014, Mar. 26, 1990; 62 FR 53538, Oct. 15, 1997]

## 21 CFR Ch. I (4–1–05 Edition)

### Subpart C—Plant and Facilities

#### § 606.40 Facilities.

Facilities shall be maintained in a clean and orderly manner, and shall be of suitable size, construction and location to facilitate adequate cleaning, maintenance and proper operations. The facilities shall:

(a) Provide adequate space for the following when applicable:

(1) Private and accurate examinations of individuals to determine their suitability as blood donors.

(2) The withdrawal of blood from donors with minimal risk of contamination, or exposure to activities and equipment unrelated to blood collection.

(3) The storage of blood or blood components pending completion of tests.

(4) The quarantine storage of blood or blood components in a designated location pending repetition of those tests that initially gave questionable serological results.

(5) The storage of finished products prior to distribution.

(6) The quarantine storage, handling and disposition of products and reagents not suitable for use.

(7) The orderly collection, processing, compatibility testing, storage and distribution of blood and blood components to prevent contamination.

(8) The adequate and proper performance of all steps in plasmapheresis, plateletpheresis and leukapheresis procedures.

(9) The orderly conduction of all packaging, labeling and other finishing operations.

(b) Provide adequate lighting, ventilation and screening of open windows and doors.

(c) Provide adequate, clean, and convenient handwashing facilities for personnel, and adequate, clean, and convenient toilet facilities for donors and personnel. Drains shall be of adequate size and, where connected directly to a sewer, shall be equipped with traps to prevent back-siphonage.

(d) Provide for safe and sanitary disposal for the following:

(1) Trash and items used during the collection, processing and compatibility testing of blood and blood components.